



Defibrillator Compatibility

Match Defibrillator Brands to Nissha Medical Part Numbers

BRAND OF DEFIBRILLATOR	MODEL OF DEFIBRILLATOR	COMPATIBLE DEFIBRILLATION ELECTRODE PART NUMBERS	
Philips Medical	<ul style="list-style-type: none"> • XLT • HeartStart XL • HP Codemaster XL (Plug Style Connector) 	<ul style="list-style-type: none"> • MRX • MRX ALS 	<ul style="list-style-type: none"> • T100AC-Philips • T100LOAC-Philips • C100AC-Philips
	<ul style="list-style-type: none"> • HeartStart FR2+ • HeartStart Forerunner 		<ul style="list-style-type: none"> • T100AC-Philips • T100LOAC-Philips
	<ul style="list-style-type: none"> • HeartStart FR3 	<ul style="list-style-type: none"> • HeartStart FRx 	NOT COMPATIBLE*
Physio-Control	<ul style="list-style-type: none"> • LP-9P • LP-10C • LP-12 	<ul style="list-style-type: none"> • LP-15 • LP-20 	<ul style="list-style-type: none"> • T100AC-Physio • T100LOAC-Physio • C100AC-Physio
	<ul style="list-style-type: none"> • LP-500 • LP-1000 	<ul style="list-style-type: none"> • LP-500 DPS 	<ul style="list-style-type: none"> • T100AC-Physio • T100LOAC-Physio
Zoll Medical	<ul style="list-style-type: none"> • M-Series • R-Series 	<ul style="list-style-type: none"> • E-Series • X-Series 	<ul style="list-style-type: none"> • T100AC-Zoll • T100LOAC-Zoll • C100AC-Zoll
	<ul style="list-style-type: none"> • Zoll AED PLUS 	<ul style="list-style-type: none"> • Zoll AED Pro 	NOT COMPATIBLE*
Cardiac Science	<ul style="list-style-type: none"> • PowerHeart AED G3 Plus • PowerHeart AED G3 Pro 		<ul style="list-style-type: none"> • T100-CS

*HeartSync pediatric defibrillation electrodes are NOT compatible with AED's. AED's cannot be manually adjusted to reduce energy delivery. Please refer to your AED's manufacturer operation manual.

Please contact Nissha Medical Technologies at 1.800.669.6905 or HS-Info@NisshaMedical.com for any questions regarding electrode compatibility,